



FLIR COMMERCIAL SYSTEMS, INC. - EXTECH BRAND
9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

Declaration of Conformity

Extech Model: 42270

Description: Humidity/Temperature Datalogge for 42275

Date of Issue: 07-Nov-13

We, FLIR Commercial Systems, Inc. - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking according to:

EMC Directive: 89/336/EEC

Report Number: G5M201020042-E-16

Report Date of Issue: 8/7/2001

Standards:

EN 55022:1999

EN 61000-4-2:1996

EN 61000-4-3:1999

The test reports show that the product fulfills the requirement in the EC EMC Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC EMC Directive.



Tony Campagna
Director of Quality Services