



FLIR COMMERCIAL SYSTEMS, INC. - EXTECH BRAND
9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

Declaration of Conformity

Extech Model: 42500
Description: IR Thermometer
Date of Issue: 07-Nov-13


We, FLIR Commercial Systems, Inc. - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking according to:

EMC Directive: 89/336/EEC
Report Number: ET92S-05-032
Report Date of Issue: 5/12/2003

Standards:

EN61326:1997/A1:1998
EN 55011:1998/A1:2000 (Group 1, Class B)
IEC 61000-4-2:1995/A1:1998/A2:2000
IEC 61000-4-3:1995/A1:1998/A2:2000

The test reports show that the product fulfills the requirement in the EC EMC Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC EMC Directive.



Tony Campagna
Director of Quality Services